



**OIML Member state**  
The Netherlands

**OIML Certificate N° R76/1992-NL1-09.15**  
Project number 810459  
Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Person responsible: Ing. C. Oosterman

### Applicant

Name: Shinko Denshi CO., Ltd.

Address: 3-9-11 Yushima, Bunkyo-ku  
Tokyo 113-0034  
Japan

### Manufacturer of the certified type

Name: Shinko Denshi CO., Ltd.

Address: 3-9-11 Yushima, Bunkyo-ku  
Tokyo 113-0034  
Japan

### Identification of certified type

Non-automatic weighing instrument  
Type: CJ

$220 \text{ g} \leq \text{Max} \leq 15 \text{ kg}$  or the equivalent in ct  
 $e \geq 0.01 \text{ g}$  or the equivalent in ct  
 $n \leq 82000$  divisions

Temperature range  $+5 \text{ }^\circ\text{C} / +35 \text{ }^\circ\text{C}$

**NMI Certin BV**  
Hugo de Grootplein 1, 3314 EG Dordrecht  
PO Box 394, 3300 AJ Dordrecht, NL  
T +31 78 6332332  
F +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision  
that no liability is accepted and that the  
applicant shall indemnify third-party liability.

Reproduction of the complete document is  
permitted.  
The notification of NMI Certin as Issuing  
Authority can be verified at [www.oiml.org](http://www.oiml.org).





**OIML Member state**  
The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number 7530 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
Edition 1992 (E)  
for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:  
N° R76/1992-NL1-09.15a that includes 48 pages;  
N° R76/1992-NL1-09.15b that includes 21 pages.

The Issuing Authority NL1  
NMI Certin, 10 April 2009

  
C. Oosterman  
Head Certification Board

\*  
\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.