

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Person responsible: Ing. C. Oosterman

Applicant

Name: Rinstrum Pty. Ltd.

Address: 41 Success Street
Acacia Ridge QLD 4110
Australia

Manufacturer of the certified type

Name: Rinstrum Pty. Ltd.

Address: 41 Success Street
Acacia Ridge QLD 4110
Australia

Identification of certified type

Non-automatic weighing instrument
Family of type: R3xx series

$100 \text{ kg} \leq \text{Max} \leq 4000 \text{ kg}$
 $e \geq 50 \text{ g}$
 $n \leq 4000 \text{ divisions}$

Temperature range $-10^{\circ}\text{C} / +40^{\circ}\text{C}$

NMI Certin BV

Hugo de Grootplein 1, 3314 EG Dordrecht
PO Box 394, 3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

Reproduction of the complete document is
permitted.
The notification of NMI Certin as Issuing
Authority can be verified at www.oiml.org.

OIML Member state
The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6244 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
Edition 1992 (E)
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° 215148a that includes 43 pages;
N° 215148b that includes 14 pages;
N° 215148c that includes 15 pages;
N° 312260 that includes 11 pages;
N° 506180 that includes 25 pages;
N° 812048 that includes 7 pages;
N° R60/2000-NL1-04.07A that includes 40 pages;
N° R60/2000-NL1-04.07B that includes 37 pages;
N° R60/2000-NL1-04.07C that includes 37 pages.

The Issuing Authority NL1
NMI Certin, 24 April 2009



C. Oosterman
Head Certification Board

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.