



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-09.27 revision 1  
Project number 9200058  
Page 1 of 2

Issuing authority	NMI Certin B.V. Person responsible: C. Oosterman
Applicant	Tanita Corporation 14-2, 1 Chome, Maeno-cho Itabash-ku Tokyo 174-8630 Japan
Manufacturer	Tanita Corporation 14-2, 1 Chome, Maeno-cho Itabash-ku Tokyo 174-8630 Japan
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : SC-240MA
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76** - Edition 1992 for accuracy class **(111)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
22 November 2010

  
C. Oosterman  
Head Certification Board

NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-09.27 revision 1  
Project number 9200058  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

Number R76/1992-NL1-09.27A, dated 1 October 2009 that includes 34 pages;  
Number R76/1992-NL1-09.27B, dated 1 October 2009 that includes 24 pages;  
Number 9200058C, dated 22 November 2010, that includes 12 pages.

**Characteristics of the non-automatic weighing instrument:**

Maximum capacity	:	Max $\leq$ 200 kg
Verification scale interval	:	e $\geq$ 100 g
Maximum number of scale intervals	:	n $\leq$ 2000 divisions
Temperature range	:	+5 °C / +35 °C