

OIML Member state
The Netherlands

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Person responsible: Ing. C. Oosterman

Applicant

Name: Mettler-Toledo (Changzhou) Measurement Technology Ltd.

Address: 111, West HaiHu Rd,
ChangZhou, XinBei District
213125 Jiangsu
P.R. of China

Manufacturer of the certified type

Name: Mettler-Toledo (Changzhou) Measurement Technology Ltd.

Address: 111, West HaiHu Rd,
ChangZhou, XinBei District
213125 Jiangsu
P.R. of China

Identification of certified type

Non-automatic weighing instrument

Type: bPro... / bC-... / RL20-...

$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

$e \geq 1 \text{ g}$

$n \leq 3000$ divisions (per weighing range) in case of multi-interval

$n \leq 6000$ divisions in case of single interval

Maximum of two partial weighing ranges

Temperature range $-10^\circ\text{C} / +40^\circ\text{C}$

In case of the instrument using scale bases covered by a Test Certificate using
the IDNet interface : $6 \text{ kg} \leq \text{Max} \leq 1500 \text{ kg}$

NMI Certin BV

Hugo de Grootplein 1, 3314 EG Dordrecht
PO Box 394, 3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

Reproduction of the complete document is
permitted.
The notification of NMI Certin as Issuing
Authority can be verified at www.oiml.org.

OIML Member state
The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6953 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
Edition 1992 (E)
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

- N° R76/1992-NL1-06.18A that includes 44 pages;
- N° R76/1992-NL1-06.18B that includes 28 pages;
- N° R76/1992-NL1-07.11A that includes 16 pages;
- N° R76/1992-NL1-07.11B that includes 26 pages;
- N° R76/1992-NL1-09.30 that includes 12 pages.

The Issuing Authority, NL1
NMI Certin, 30 October 2009



C. Oosterman
Head Certification Board

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.