

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Shanghai Teraoka Electronic Co.,LTD.
Address: Tinglin Industry Development Zone
201505 Shanghai
PR of China

Manufacturer of the certified type

Name: Shanghai Teraoka Electronic Co.,LTD.
Address: Tinglin Industry Development Zone
201505 Shanghai
PR of China

Identification of certified type

Non-automatic weighing instrument
Type : SM-100., SM-5100..

$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
 $e \geq 1 \text{ g}$

$n \leq 6000$ divisions for a single interval scale
 $n \leq 3000$ divisions (per partial weighing range)

maximum of 2 partial weighing ranges

$T \leq -\frac{1}{2} \text{ Max}$ for single-interval scale

$T \leq -\text{Max}_1$ for multi-interval scale

Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6734 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
Edition 1992 (E)
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

- N° R76/1992-NL1-05.11 that includes 57 pages;
- N° R76/1992-NL1-05.18 that includes 15 pages;
- N° R76/1992-NL1-04.13A that includes 56 pages;
- N° R76/1992-NL1-05.31A that includes 24 pages;
- N° R76/1992-NL1-05.31B that includes 16 pages;
- N° R76/1992-NL1-05.40 that includes 16 pages;
- N° R76/1992-NL1-07.03 that includes 43 pages;
- N° R76/1992-NL1-09.22A that includes 15 pages;
- N° R76/1992-NL1-09.22B that includes 13 pages;
- N° R76/1992-NL1-09.33 that includes 13 pages.

The Issuing Authority NL1
NMI Certin, 14 January 2010



C. Oosterman
Head Certification Board

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.