



OIML Member state
The Netherlands

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Person responsible: Ing. C. Oosterman

Applicant

Name: Shinko Denshi CO., Ltd.

Address: 3-9-11 Yushima, Bunkyo-ku
Tokyo 113-0034
Japan

Manufacturer of the certified type

Name: Shinko Denshi CO., Ltd.

Address: 3-9-11 Yushima, Bunkyo-ku
Tokyo 113-0034
Japan

Identification of certified type

Non-automatic weighing instrument
Type: CJ

For type CJ class II:

$220 \text{ g} \leq \text{Max} \leq 15 \text{ kg}$ or the equivalent in ct
 $e \geq 0.01 \text{ g}$ or the equivalent in ct
 $n \leq 82000$ divisions

Temperature range $+5 \text{ }^\circ\text{C} / +35 \text{ }^\circ\text{C}$

NMI Certin BV
Hugo de Grootplein 1, 3314 EG Dordrecht
PO Box 394, 3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

Reproduction of the complete document is
permitted.
The notification of NMI Certin as Issuing
Authority can be verified at www.oiml.org.



OIML Member state

The Netherlands

For type CJ class III:

 $300 \text{ g} \leq \text{Max} \leq 12 \text{ kg}$ or the equivalent in ct $e \geq 0.1 \text{ g}$ or the equivalent in ct $n \leq 6000$ divisionsTemperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7530 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class $\textcircled{\text{II}}$ or $\textcircled{\text{III}}$

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-09.15A revision 2 that includes 49 pages;

N° R76/1992-NL1-09.15B revision 2 that includes 21 pages;

N° R76/1992-NL1-10.04A that includes 32 pages;

N° R76/1992-NL1-10.04B that includes 23 pages.

The Issuing Authority NL1
NMI Certin, 03 March 2010



C. Oosterman
Head Certification Board

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.