

## IML Certificate Conformity

**OIML Member State** The Netherlands

Number R76/1992-NL1-10.18 revision 1 Project number 10200053 Page 1 of 2

Issuing authority

NMi Certin B.V.

Person responsible: C. Oosterman

Applicant

**Tanita Corporation** 

14-2,1 chome, Maeno-cho

Itabashi-ku

Tokyo, 174-8630

Manufacturer

**Tanita Corporation** 

14-2,1 chome, Maeno-cho

Itabashi-ku Tokyo, 174-8630

Identification of the

A Non-automatic weighing instrument

certified type

Type

MC-980A or MC-980MA

Characteristics

See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 1992 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority

30 November 2010

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







## OIML Certificate of Conformity

**OIML Member State** The Netherlands

Number R76/1992-NL1-10.18 revision 1 Project number 10200053 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

Number R76/1992-NL1-10.18A revision 1 dated 13 December 2010 that includes 48 pages; Number R76/1992-NL1-10.18B revision 1 dated 13 December 2010 that includes 8 pages.

## Characteristics of the non-automatic weighing instrument:

 $\label{eq:max} \begin{tabular}{lll} Maximum capacity & : & Max \le 300 \ kg \\ \begin{tabular}{lll} Verification scale interval & : & e \ge 0.1 \ kg \\ \end{tabular}$ 

Maximum number of scale intervals :  $n \le 3000$  divisions Temperature range : +5 °C / +35 °C

Tare : Only preset tare existent