

OIML Member State
The Netherlands

Number R76/1992-NL1-11.22
Project number 11200546
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Teraoka Seiko Co. Ltd. 13-12 Kugahara 5-Chome Ohta-Ku Tokyo 146-8580 Japan
Manufacturer	Teraoka Seiko Co. Ltd. 13-12 Kugahara 5-Chome Ohta-Ku Tokyo 146-8580 Japan
Identification of the certified type	A Non-automatic weighing instrument Type : DS-805..
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 1992 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
23 September 2011


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/1992-NL1-11.22
Project number 11200546
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

Number NMI-11200546-01 dated 23 September 2011 that includes 51 pages;
Number NMI-11200546-02 dated 23 September 2011 that includes 25 pages.

Characteristics of the non-automatic weighing instrument:

- Maximum capacity : $3 \text{ kg} \leq \text{Max} \leq 300 \text{ kg}$
- Verification scale interval : $e \geq 1 \text{ g}$
- Maximum number of scale intervals : $n \leq 3000$ divisions
(per partial weighing range)
- Maximum partial weighing ranges : 2
- Temperature range : $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
- Tare : $T \leq -\text{Max}_1 - e_1$ for multi interval instruments