



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/1992-NL1-13.10
Project number 12200108
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Teraoka Seiko Co., Ltd. 13-12 Kugahara 5-Chome, Ohta-Ku, Tokyo 146-8580, Japan
Manufacturer	Teraoka Seiko Co., Ltd. 13-12 Kugahara 5-Chome, Ohta-Ku, Tokyo 146-8580, Japan
Identification of the certified type	A Non-automatic weighing instrument Type : DPS-5600
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 1992 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
13 May 2013

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/1992-NL1-13.10
Project number 12200108
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- Number 209887 dated 5 June 2003 that includes 55 pages;
- Number R76/1992-NL1-07.10 dated 1 March 2007 that includes 15 pages;
- Number R76/1992-NL1-07.24 dated 9 July 2007 that includes 8 pages;
- Number R76/1992-NL1-10.10 dated 16 March 2010 that includes 34 pages;
- Number R76/1992-NL1-10.11 dated 16 March 2010 that includes 26 pages.
- Number NMI-12200108-02 dated 10 October 2012 that includes 36 pages;
- Number NMI-12200108-03 dated 10 October 2012 that includes 18 pages;
- Number NMI-12200108-04 dated 10 October 2012 that includes 18 pages;
- Number NMI-12200108-05 dated 10 October 2012 that includes 18 pages;
- Number NMI-12200108-06 dated 10 October 2012 that includes 18 pages;
- Number NMI-12200108-07 dated 10 October 2012 that includes 18 pages;
- Number NMI-12200108-12 dated 13 May 2013 that includes 9 pages;
- Number NMI-12200108-13 dated 13 May 2013 that includes 15 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 150 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Maximum number of scale intervals	$n \leq 3000$ divisions (per partial weighing range)
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Tare	$T \leq -\text{Max}$ for single interval instruments $T \leq -\text{Max}_1$ for multi interval instruments
Weighing range(s)	Single interval Multi-interval
Maximum partial weighing ranges	2
Application	Intended to be used as price labeling instrument
Power supply voltage	100 –240 V AC 50/60 Hz