

OIML Member State
The Netherlands

Number R76/1992-NL1-14.29
Project number 14200224
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Snowrex International Co., Ltd. 2F, No. 9, Lane 50, Sec. 3 Nan Kang Road, Taipei Taiwan R.O.C.
Identification of the certified type	A Non-automatic weighing instrument Type : PS-series
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 1992 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
11 July 2014



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/1992-NL1-14.29
Project number 14200224
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. 208303 dated 11 February 2003 that includes 45 pages;
- No. R60/1991-NL-95.11 dated 10 July 1995 that includes 38 pages (HBM SP4);
- No. R60/1991-NL-95.15 dated 13 October 1995 that includes 35 pages (HBM SP4);
- No. R60/1991-NL-98.13 that includes 12 pages (HBM SP4);
- No. R60/1991-NL-98.07 dated 6 April 1998 that includes 39 pages (HBM PW12);
- No. R60/1991-NL-97.25 dated 24 September 1997 that includes 38 pages (Minebea CB14);
- No. R60/1991-NL-97.26 dated 24 September 1997 that includes 38 pages (Minebea U2D1-F);
- No. R60/1991-NL-98.15A dated 24 September 1998 that includes 35 pages (Minebea CB004-NS);
- No. R60/1991-NL-98.15B dated 24 September 1998 that includes 38 pages (Minebea CB004-NS);
- No. R60/1991-NL-98.20 dated 19 November 1998 that includes 37 pages (Minebea C2G1-F);
- No. R60/1991-NL-98.23 dated 7 December 1998 that includes 35 pages (Minebea C2G1-F);
- No. NMI-14200224-01 dated 11 July 2014 that includes 11 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 3000 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing range(s)	Single interval
Maximum number of scale intervals (one weighing range)	$n \leq 3000$ divisions
Temperature range	$0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	9 V DC by an external power supply, 6 V DC by an internal battery.
Software identification	Version number: PSP N1.