

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-14.56  
Project number SO14204085  
Page 1 of 2

|                                      |   |
|--------------------------------------|---|
| Issuing authority                    | NMi Certin B.V.<br>Person responsible: C. Oosterman   |
| Applicant and Manufacturer           | Mettler Toledo (Changzhou) Measurement Technology Ltd.<br>No.111, West TaiHu Road<br>ChangZhou XinBei District<br>213125 Jiangsu<br>Peoples Republic of China |
| Identification of the certified type | <b>A Non-automatic weighing instrument</b><br>Type : RN20 / VIVA  |
| Characteristics                      | See next page   |

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
21 October 2014

C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-14.56  
Project number SO14204085  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-03.21 that includes 56 pages;
- No. R76/1992-NL1-04.34 that includes 18 pages;
- No. R76/1992-NL1-07.13 that includes 36 pages.

## Characteristics of the non-automatic weighing instrument:

|                                   |  |
|-----------------------------------|--|
| Accuracy class                    | III  |
| Maximum capacity                  | $3 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$ or $6 \text{ lb} \leq \text{Max} \leq 30 \text{ lb}$ |
| Verification scale interval       | $e \geq 1 \text{ g}$ or $e \geq 0,002 \text{ lb}$  |
| Maximum number of scale intervals | $n \leq 3000$ divisions<br>(per partial weighing range)  |
| Tare                              | $T \leq -\text{Max}$   |
| Weighing range(s)                 | Single interval  |
| Temperature range                 | $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$  |
| Power supply voltage              | 9-12 V AC/DC adapter   |
| Application                       | Intended to be used for direct sales to the public   |