



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-13.17 Revision 1  
Project number 1901500  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	CAS Corporation #262, Geurugogae-ro, Gwangjeok-myeon, Yangju-si, Gyeonggi-do, Rep. of Korea
Manufacturer	CAS Corporation #262, Geurugogae-ro, Gwangjeok-myeon, Yangju-si, Gyeonggi-do, Rep. of Korea or, CAS (Zhejiang) Electronics Co., Ltd. 99# Changjiang Road, Jiashan County, Zhejiang Province, China
	An <b>Indicator</b> Type : CI-200D, CI-201D, CI-200SD, CI-200SCD, CI-201SD
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76** - Edition 2006 for accuracy class **(III)** and **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
18 September 2017

  
C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-13.17 Revision 1  
Project number 1901500  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-12200013-01 dated 24 May 2013 that includes 46 pages;
- No. NMI-12200013-02 dated 24 May 2013 that includes 16 pages.

## Characteristics of the indicator:

Accuracy class	Ⓜ and ⓂⓂ
Temperature range	-10 °C / +40 °C
Fraction of the maximum permissible error	0
Weighing range	Single interval
Power supply voltage	110-240 V AC 50/60 Hz
Maximum number of digital load cells	8
Software identification	Version number: U 2.04
Power supply to digital load cell	9 V DC

## Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
Initial	31 May 2013	-
1	18 September 2017	CAS Corporation address revised, CAS (Zhejiang) Electronics Co., Ltd. added as additional manufacturer.