



Member State
Switzerland

OIML Certificate
No R76_2006-CH1-08.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name Federal Office of Metrology METAS
Certification Body METAS-Cert

Address METAS, Lindenweg 50, CH-3003 Bern-Wabern

Person responsible Jürg Ramseyer, Head of METAS-Cert

Applicant

Name Bühler AG

Address 9240 Uzwil, Schweiz

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified pattern

Non-automatic electromechanical weighing instrument

Type **MEAG**

Specifications See on page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 76-1, edition 2006

for accuracy class(es) **III**, **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



**OIML Certificate
No R76_2006-CH1-08.01**

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

No 513-00003, including 43 pages

Specifications:

Accuracy class	III	III
Max	10 kg to 99 000 kg	
e	1 g to 50 kg	
n	≤ 5 000	≤ 1 000
Tare-balancing range	Until 100 % of Max	
Preset tare value	Until 100 % of Max	
Temperature range	-10 °C / 50 °C	
Power supply	95 VAC to 250 VAC or 24 VDC	
LC-impedance	43 Ω to 350 Ω	
Minimum Load input signal	0.6 μV/e	

The Issuing Authority

The OIML Member

Federal Office of Metrology METAS

Jürg Ramseyer, Head of METAS-Cert

Dr. Philippe Richard, Vice Director

CH-3003 Bern-Wabern, 15 December 2008

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and the associated Test Report is not permitted, although either may be reproduced in full.