



**Member State**  
**Switzerland**

**OIML Certificate**  
**No R76/2006-CH1-08.02**

## OIML CERTIFICATE OF CONFORMITY

### *Issuing authority*

<i>Name</i>	Federal Office of Metrology METAS Certification Body METAS-Cert
<i>Address</i>	METAS, Lindenweg 50, CH-3003 Bern-Wabern
<i>Person responsible</i>	Jürg Ramseyer, Head of METAS-Cert

### *Applicant*

<i>Name</i>	Haenni Wheel Load Scales AG
<i>Address</i>	3303 Jegenstorf, Schweiz
<i>Manufacturer</i>	The manufacturer of the certified pattern is the Applicant

### *Identification of the certified pattern*

**Non-automatic electromechanic wheel or axle weighing instrument**

<i>Type</i>	<b>WL 104</b>
<i>Max</i>	≤ 10 000 kg
<i>e</i>	≥ 50 kg
<i>n</i>	200
<i>Temperature range</i>	-20 °C / +60 °C

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 76-1, edition 2006  
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



**OIML Certificate  
No R76/2006-CH1-08.02**

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

No 513-00019, including 56 pages

*Specifications:*

Accuracy class	Ⓜ
Max	≤ 10 000 kg
e	≥ 50 kg
n	≤ 200
Temperature range	-20 °C / +60 °C
Power supply	11.5 V DC / 16 V DC

**The Issuing Authority**

Federal Office of Metrology METAS

P.P.

Jürg Ramseyer, Head of METAS-Cert

**The OIML Member**

Dr. Philippe Richard, Vice Director

CH-3003 Bern-Wabern, 19 December 2008

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and the associated Test Report is not permitted, although either may be reproduced in full.