



Member State
Switzerland

OIML Certificate
No R76_2006-CH1-09.01 Rev. 1

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name **Federal Office of Metrology METAS**
Certification Body METAS-Cert

Address Lindenweg 50
CH-3003 Bern-Wabern

Person responsible Jürg Ramseyer, Head of METAS-Cert

Applicant

Name **METTLER-TOLEDO AG**
LabTec

Address Im Langacher
8606 Greifensee
Switzerland

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified pattern

Non-automatic weighing instrument

Type **NewClassic MF**

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



R 76-1, edition 2006

for accuracy class(es) **I**, **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



**OIML Certificate
No R76_2006-CH1-09.01 Rev. 1**

Metrological characteristics

Accuracy class	I	II	
		Small housing	Large housing
Max	≤ 10 200 g	120 g to 10 200 g	10 200 g to 64 200 g
e	≥ 1 mg	≥ 10 mg	≥ 1 g
d	≥ 0.01 mg	≥ 1 mg	≥ 0.1 g
n	≤ 420 000	≤ 82 000	≤ 64 200
Tare-balancing range	Until 100 % of Max		
Temperature range	10 °C / 30 °C		5 °C / 40 °C

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

No	Including pages	No	Including pages
511-00144	28	511-00354	28
511-00152	32	511-00357	32
511-00306	32	511-00361	28
511-00339	18	511-00441	22
511-00342	28	511-00523	24
511-00345	28	511-00524	17
511-00348	20	511-00525	17
511-00351	17	511-00527	46
		511-00530	28

The Issuing Authority

Federal Office of Metrology METAS

Jürg Ramseyer, Head of METAS-Cert

The OIML Member

Dr. Philippe Richard, Vice Director

CH-3003 Bern-Wabern, 3 May 2010

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and the associated Test Report is not permitted, although either may be reproduced in full.