



Member State
Switzerland

OIML Certificate
No R76/2006-CH1-12.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name **Federal Office of Metrology METAS**
Certification Body METAS-Cert

Address Lindenweg 50
CH-3003 Bern-Wabern

Person responsible Gulian Couvreur, Head of METAS-Cert

Applicant

Name **Mettler-Toledo AG**

Address Im Langacher 44
8606 Greifensee
Schweiz

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified pattern

Non-Automatic Weighing Instrument

Type **ME ... / TLE ... / JE ... / PHE ... (E-Line)**

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



R 76-1, edition 2006

for accuracy class(es) (I), (II)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



Metrological characteristics

Accuracy class	Ⓘ	Ⓜ
Max	≤ 450 g	≤ 4 500 g
e	≥ 0.001 g	≥ 0.01 g
n	≤ 250 000	≤ 45 000
Tare-balancing range	Until 100 % of Max	
Temperature range	10 °C / 30 °C	

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

No	Including pages
511-01039	30
511-01040	41
511-01041	41
511-01042	23

CH-3003 Bern-Wabern, 19 September 2012

The Issuing Authority

Federal Office of Metrology METAS

Gullian Couvreur, Head of METAS-Cert

The OIML Member

Dr. Philippe Richard, Deputy Director

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and the associated Test Report is not permitted, although either may be reproduced in full.