



**Member State**  
**Switzerland**

**OIML Certificate**  
**No R76/2006-CH1-16.01 Rev. 1**

## OIML CERTIFICATE OF CONFORMITY

*Issuing authority*

*Name* **Federal Institute of Metrology METAS**  
Conformity Evaluation Body METAS-Cert

*Address* Lindenweg 50, 3003 Bern-Wabern, Switzerland

*Person responsible* Gulian Couvreur, Head of METAS-Cert

*Applicant*

*Name* Mettler-Toledo GmbH

*Address* Im Langacher 44, 8606 Greifensee, Switzerland

*Manufacturer* The manufacturer of the certified pattern is the Applicant

*Identification of the certified pattern*

**Non-automatic analytical precision weighing instrument**

*Type* **ME...T**

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



**R 76-1, edition 2006**

for accuracy class(es) **I**, **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

**OIML Certificate**  
**No R76/2006-CH1-16.01 Rev. 1**

Accuracy class	Ⓘ	Ⓜ
Max	≤ 220 g	≤ 5 200 g
e	≥ 1 mg	≥ 10 mg
n	≤ 220 000	≤ 52 000
Tare-weighing range	Until 100 % of Max	
Temperature range	10 °C / 30 °C	

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

No / Including pages

6030-00039 / 38

6030-00040 / 49

6030-00041 / 49

6030-00044 / 23

3003 Bern-Wabern, 10 August 2016

**The Issuing Authority**



Gulian Couvreur, Head of METAS-Cert

**The OIML Member**



Dr. Gregor Dudle, Deputy Director

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and the associated Test Report is not permitted, although either may be reproduced in full.