



Member State
Switzerland

OIML Certificate No.
R076/2006-A-CH1-2018.02 Rev. 1.

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

<i>Name</i>	Federal Institute of Metrology METAS Conformity Evaluation Body METAS-Cert
<i>Address</i>	Lindenweg 50, 3003 Bern-Wabern, Switzerland
<i>Person responsible</i>	Gulian Couvreur, Head of METAS-Cert

Applicant

<i>Name</i>	Mettler-Toledo GmbH
<i>Address</i>	Im Langacher 44, 8606 Greifensee, Schweiz

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified type

XPR / XSR Analytical Balances

Type **XPR / XSR**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R 76-1, edition 2006

for accuracy class **1**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Annex to the OIML-CS Certificate No R076/2006-A-CH1-2018.02 Revision 1

Metrological characteristics

Accuracy class	Ⓛ
Max	22 g to 320 g
d	≥ 1 µg
e	1 mg
n	≤ 320.000
Tare-balancing range	Until 100 % of Max
Preset tare value	Until 100 % of Max
Temperature range	10 °C / 30 °C

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
6030-01348	26.10.2020	10

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
CH-W1-18026-01_XPR_XSR_ Doc_List_TEC	26.10.2020	5

OIML Certificate History:

Revision No.	Date	Description of the modification
00	28.08.2018	First issue
01	26.10.2020	Editorial changes

The OIML Issuing Authority CH1

3003 Bern-Wabern, 26.10.2020

Approved by

Gulian Couvreur, Head of sector
METAS-Cert

Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full