



Member State
Switzerland

OIML Certificate No.
R076/2006-A-CH1-2018.06

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

<i>Name</i>	Federal Institute of Metrology METAS Conformity Evaluation Body METAS-Cert
<i>Address</i>	Lindenweg 50, 3003 Bern-Wabern
<i>Person responsible</i>	Gulian Couvreur, Head of METAS-Cert

Applicant

<i>Name</i>	Mettler-Toledo Instruments
<i>Address</i>	Gui Ping Road 589, 200233 Shanghai, Peoples Republic of China

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified type

Non-Automatic Weighing Instrument

Type **ME ... / TLE ... / JE ... / PHE ...**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R 76-1, edition 2006

for accuracy class(es) (I), (II)

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Annex to OIML-CS Certificate No R076/2006-A-CH1-2018.06

Metrological characteristics

Accuracy class	Ⓛ	Ⓜ
Max	≤ 250 g	≤ 5 200 g
e	≥ 0.001 g	≥ 0.01 g
n	≤ 250 000	≤ 52 000
Tare-balancing range	Until 100 % of Max	
Temperature range	10 °C / 30 °C	

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
511-01039	13.09.2012	30
511-01040	13.09.2012	41
511-01041	13.09.2012	41
511-01042	13.09.2012	23
6030-00337	07.12.2016	46
6030-00338	07.12.2016	11
6030-00339	07.12.2016	49
133-08231	17.07.2018	46

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
CH-W1-12013-01_ME_TLE_JE_PHE_DocList	18.07.2018	4

OIML Certificate History:

Revision No.	Date	Description of the modification
00	26.07.2018	Revision of R76/2006-CH1 -16.04 including new test report

The OIML Issuing Authority CH1

3003 Berne-Wabern, 26 July 2018

Approved by

Gulian Couvreur, Head of sector
METAS-Cert



Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full