



**Member State**  
**Switzerland**

**OIML Certificate No.**  
**No R076/2006-A-CH1-2019.02 Rev. 1.**

## OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

### *Issuing authority*

*Name* **Federal Institute of Metrology METAS**  
Conformity Evaluation Body METAS-Cert

*Address* **Lindenweg 50, 3003 Bern-Wabern, Switzerland**

*Person responsible* **Gulian Couvreur, Head of METAS-Cert**

### *Applicant*

*Name* **Sartorius Lab Instruments GmbH & Co KG**

*Address* **Otto-Brenner-Strasse 20, 37079 Göttingen, Germany**

*Manufacturer* **The manufacturer of the certified pattern is the applicant.**

### *Identification of the certified type*

**Precision balances BC-E and BC-A Series**

### *Type*

**BC-EI, BC-EJ, BC-AI, BC-AJ**

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



**R 76-1, edition 2006**

for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

# Annex to the OIML-CS Certificate No R076/2006-A-CH1-2019.02 Revision 1

## Metrological characteristics

Type	BC-EI, BC-AI	BC-EJ, BC-AJ
Accuracy Class	(II)	
Max	5 000 g ... 8 200 g	5 000 g ... 8 200 g
e	1 g	1 g
d	0.1 g - 1 g	0.1 g - 1 g
n	≤ 8 200	≤ 8 200
Tare-balancing range	until 100% of Max	
Temperature range	+10 °C / +30 °C	
Nominal capacity of the load receptor	9 840 g	9 840 g
Initial zero setting + dead load <sup>1)</sup>	≤ 4 840 g	≤ 4 840 g
Maximum weighing pan size	180 mm x 180 mm	Ø 180 mm

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
6030-1190-01	2019-09-24	12

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
R076_2006-A-CH1-2019.02-01_SARTORIUS_LERD	2020-09-24	4

## Annex to the OIML-CS Certificate No R076/2006-A-CH1-2019.02 Revision 1

### *OIML Certificate History:*

Revision No.	Date	Description of the modification
00	2019-09-24	First issue
01	2020-09-24	Extension with the types BC-AI, BC-AJ

### **The OIML Issuing Authority CH1**

3003 Bern-Wabern, 24 September 2020

*Approved by*

Gulian Couvreur, Head of sector  
METAS-Cert



### *Important note:*

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full