



**Member State**  
**Switzerland**

**OIML Certificate**  
**No R076/2006-A-CH1-2019.02**

## OIML CERTIFICATE OF CONFORMITY

### *Issuing authority*

*Name* **Federal Institute of Metrology METAS**

*Address* **Lindenweg 50, 3003 Bern-Wabern, Switzerland**

*Person responsible* **Gulian Couvreur, Head of METAS-Cert**

### *Applicant*

*Name* **Sartorius Lab Instruments GmbH & Co KG**

*Address* **Otto-Brenner-Strasse 20, 37079 Göttingen, Germany**

*Manufacturer* **The manufacturer of the certified pattern is the applicant.**

### *Identification of the certified pattern*

**Precision balances BC-E Line**

*Type* **BC-EI, BC-EJ**

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



**R 76-1, edition 2006**

for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

*Metrological characteristics*

Type	BC-EI	BC-EJ
Accuracy Class	(II)	
Max	5000 g ... 8200 g	5000 g ... 8200 g
e	1 g	1 g
d	0.1 g ... 1 g	0.1 g ... 1 g
n	≤ 8200	≤ 8200
Tare-balancing range	until 100% of Max	
Temperature range	+10 °C / +30 °C	
Nominal capacity of the load receptor	9840 g	9840 g
Initial zero setting + dead load <sup>1)</sup>	≤ 4840 g	≤ 4840 g
Maximum weighing pan size	180 mm x 180 mm	Ø 180 mm

*The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:*

No.	Date	Including pages
133-08524	20.08.2019	62
6030-1190	23.09.2019	10

*The technical documentation relating to the identified type is contained in documentation file:*

Name	Date	Including pages
CH-W1-19031-00_DE_SARTORIUS_Doc _List_TEC	24.09.2019	3

*OIML Certificate History:*

Revision No.	Date	Description of the modification
00	24.09.2019	First issue

**The OIML Issuing Authority CH1**

3003 Bern-Wabern, 24 September 2019

*Approved by*

Gulian Couvreur, Head of sector  
METAS-Cert



*Important note:*

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full