



**Member State**  
**Switzerland**

**OIML Certificate No.**  
**R076/2006-A-CH1-2019.03**

## OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

### *Issuing authority*

*Name* **Federal Institute of Metrology METAS**  
Conformity Evaluation Body METAS-Cert

*Address* Lindenweg 50, 3003 Bern-Wabern, Switzerland

*Person responsible* Gulian Couvreur, Head of METAS-Cert

### *Applicant*

*Name* **Mettler-Toledo AG**

*Address* Im Langacher 44, 8606 Greifensee, Switzerland

*Manufacturer* The manufacturer of the certified pattern is the Applicant

### *Identification of the certified type*

*Type* **Non-automatic analytical precision weighing instrument**  
**ME...T / JET**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



**OIML R 76-1, edition 2006**

for accuracy classes **I**, **II**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

## OIML Certificate No R076/2006-A-CH1-2019.03

### Metrological characteristics

#### ME..T instruments

Accuracy class	Ⓘ	Ⓜ
Max	≤ 220 g	≤ 5 200 g
e	≥ 1 g	≥ 10 g
n	≤ 220 000	≤ 52 000
Tare-weighing range	Until 100 % of Max	
Temperature range	10 °C / 30 °C	

#### JET instruments

Accuracy class	Ⓘ	Ⓜ
Max	≤ 12 200 g	≤ 6 200 g
e	≥ 1mg	≥ 10 mg
n	≤ 160 000	≤ 62 000
Tare-weighing range	Until 100 % of Max	
Temperature range	10 °C / 30 °C	

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

<b>No.</b>	<b>Date</b>	<b>Including pages</b>
6030-00039	11.07.2016	38
6030-00040	11.07.2016	49
6030-00041	11.07.2016	49
6030-00044	12.07.2016	23
133-08543	03.09.2019	57
133-08544	03.09.2019	57
6030-01218	27.09.2019	11

The technical documentation relating to the identified type is contained in documentation file:

<b>Name</b>	<b>Date</b>	<b>Including pages</b>
CH_MT_20190902_ME-T-JET_CH-W1-16021_Doc_List_TEC	27.09.2019	4

## OIML Certificate No R076/2006-A-CH1-2019.03

### *OIML Certificate History:*

Revision No.	Date	Description of the modification
00	03.09.2019	First issue, revision of R76/2006-CH1-16.01

### **The OIML Issuing Authority CH1**

3003 Bern-Wabern, 27 September 2019

*Approved by*

Gulian Couvreur, Head of sector  
METAS-Cert



*Important note:*

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full