



Member State
Switzerland

OIML Certificate No.
R076/2006-A-CH1-2019.04 Rev. 1.

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

Name **Federal Institute of Metrology METAS**
Conformity Evaluation Body METAS-Cert

Address Lindenweg 50, 3003 Bern-Wabern, Switzerland

Person responsible Gulian Couvreur, Head of METAS-Cert

Applicant

Name **Sartorius Lab Instruments GmbH & Co KG**

Address Otto-Brenner-Strasse 20, 37079 Göttingen, Germany

Manufacturer

Name **Sartorius Scientific Instruments (Beijing) Co, Limited**

Address Yu An Road No. 33, Zone B
Tianzhu Airport Industrialzone, Shunyi District,
101 300 Beijing
PR China

Identification of the certified type

Type **Precision and analysis balances BC-E and BC-A Line**

Module **BC-EA to BC-EH, BC-AA to BC-AH**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R 76-1, edition 2006

for accuracy classes (I), (II)

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Annex to OIML-CS Certificate No R076/2006-A-CH1-2019.04

Metrological characteristics

Type	BC-EA BC-AA	BC-EB BC-AB	BC-EB BC-AB
Accuracy Class	I	I	II
Max	50 g - 320 g	50 g - 220 g	0,1 g - 100 g
e	1 mg - 2 mg	1 mg - 2 mg	1 mg - 2 mg
d	0.1 mg - 2 mg	0.1 mg - 2 mg	0.1 mg - 2 mg
n	≤ 320 000	≤ 220 000	≤ 100 000
Tare-balancing range	until 100% of Max		
Temperature range 1	+17 °C / +27 °C	+17 °C / +27 °C	+10 °C / +30 °C
Temperature range 2 ¹⁾	+10 °C / +30 °C	+10 °C / +30 °C	+10 °C / +30 °C
Nominal capacity of the load receptor	384 g	264 g	264 g
Initial zero setting + dead load ²⁾	≤ 234 g	≤ 214 g	≤ 263 g
Maximum weighing pan size	Ø 90 mm	Ø 90 mm	Ø 90 mm

Type	BC-EC BC-AC	BC-ED BC-AD
Accuracy Class	I	II
Max	500 g – 1 500 g	1 g - 650 g
e	10 mg - 20 mg	0.01 g – 0.1 g
d	1 mg - 20 mg	0.001 g – 0.1 g
n	≤ 150 000	≤ 65 000
Tare-balancing range	until 100% of Max	
Temperature range 1	+17 °C / +27 °C	+10 °C / +30 °C
Temperature range 2 ¹⁾	+10 °C / +30 °C	+10 °C / +30 °C
Nominal capacity of the load receptor	1 800 g	780 g
Initial zero setting + dead load ²⁾	≤ 1300 g	≤ 779 g
Maximum weighing pan size	Ø 120 mm	Ø 120 mm

Annex to OIML-CS Certificate No R076/2006-A-CH1-2019.04

Type	BC-EE BC-AE	BC-EF BC-AF
Accuracy Class	(II)	
Max	500 g - 6 200 g	500 g - 6 200 g
e	0.1 g - 1 g	0.1 g - 1 g
d	0.01 g - 1 g	0.01 g - 1 g
n	≤ 62 000	≤ 62 000
Tare-balancing range	until 100% of Max	
Temperature range 1	+10 °C / +30 °C	
Nominal capacity of the load receptor	7 440 g	7 440 g
Initial zero setting + dead load ²⁾	6 940 g	6 940 g
Maximum weighing pan size	180 mm x 180 mm	Ø 180 mm

Type	BC-EG BC-AG	BC-EH BC-AH
Accuracy Class	(II)	
Max	5 000 g - 12 200 g	5 000 g - 12 200 g
e	1 g	1 g
d	0.1 g - 1 g	0.1 g - 1 g
n	≤ 12 200	≤ 12 200
Tare-balancing range	until 100% of Max	
Temperature range	+10 °C / +30 °C	
Nominal capacity of the load receptor	14 640 g	14 640 g
Initial zero setting + dead load ²⁾	≤ 9 640 g	≤ 9 640 g
Maximum weighing pan size	180 mm x 180 mm	Ø 180 mm

¹⁾ Only for weighing instruments with incorporated span adjustment device being automatically released.

²⁾ The sum of Max, initial zero setting range and dead load shall not exceed the nominal load of the load receptor.

Annex to OIML-CS Certificate No R076/2006-A-CH1-2019.04

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
6030-01225	2019-09-24	13
6030-01225-01	2020-09-21	19

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
R076_2006-A-CH1-2019.04_Rev_01_LERD	2020-09-23	6

OIML Certificate History:

Revision No.	Date	Description of the modification
00	2019-09-24	First issue
01	2020-09-21	Extension to the types BC-EA to BC-EH, BC-AA to BC-AH

The OIML Issuing Authority CH1

3003 Bern-Wabern, 21 September 2020

Approved by

Gulian Couvreur, Head of sector
METAS-Cert



Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full