



Member State
Switzerland

OIML Certificate No.
No R076/2006-A-CH1-2019.05 Rev. 3.

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

Name **Federal Institute of Metrology METAS**
Conformity Evaluation Body METAS-Cert

Address Lindenweg 50, 3003 Bern-Wabern, Switzerland

Person responsible Gulian Couvreur, Head of METAS-Cert

Applicant

Name **Sartorius Lab Instruments GmbH & Co KG**

Address Otto-Brenner-Strasse 20, 37079 Göttingen, Germany

Manufacturer

Name **Sartorius Scientific Instruments (Beijing) Co, Limited**

Address Yu An Road No. 33, Zone B
Tianzhu Airport Industrialzone, Shunyi District,
101 300 Beijing
PR China

Identification of the certified type

Precision balances BC-E and BC-A Line

Type

BC-EI, BC-EJ, BC-AI, BC-AJ

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



R 76-1, edition 2006

for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Annex to the OIML-CS Certificate No R076/2006-A-CH1-2019.05 Revision 3

Metrological characteristics

Type	BC-EI, BC-AI	BC-EJ, BC-AJ
Accuracy Class	(II)	
Max	5 000 g ... 8 200 g	5 000 g ... 8 200 g
e	1 g	1 g
d	0.1 g - 1 g	0.1 g - 1 g
n	≤ 8 200	≤ 8 200
Tare-balancing range	until 100% of Max	
Temperature range	+10 °C / +30 °C	
Nominal capacity of the load receptor	9 840 g	9 840 g
Initial zero setting + dead load ¹⁾	≤ 4 840 g	≤ 4 840 g
Maximum weighing pan size	180 mm x 180 mm	Ø 180 mm

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
6030-01226-03	2021-04-15	12

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
R076_2006-A-CH1- 2019.05_Rev_03_SARTORIUS_LERD_V01	2021-04-15	5

Annex to the OIML-CS Certificate No R076/2006-A-CH1-2019.05 Revision 3

OIML Certificate History:

Revision No.	Date	Description of the modification
00	2019-09-24	First issue
01	2020-09-24	Extension with the types BC-AI, BC-AJ
02	2020-11-03	Editorial changes
03	2021-04-15	Rebranding, new design

The OIML Issuing Authority CH1

3003 Bern-Wabern, 15.04.2021

Approved by

Gulian Couvreur, Head of sector
METAS-Cert



Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full