



Member State
Switzerland

OIML Certificate No.
R076/2006-A-CH1-2019.06 Rev. 1.

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

Name **Federal Institute of Metrology METAS**
Conformity Evaluation Body METAS-Cert

Address Lindenweg 50, 3003 Bern-Wabern, Switzerland

Person responsible Gulian Couvreur, Head of METAS-Cert

Applicant

Name **Mettler-Toledo Instruments (Shanghai) Co. Ltd.**

Address Gui Ping Road 589, 200233 Shanghai, People's Republic of China

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified type

Type **Non-automatic analytical precision weighing instrument**
ME...T / JET

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R 76-1, edition 2006

for accuracy classes ①, ②

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The present revision of this certificate is the only valid and replaces all previous revisions.
This document is only valid and reviewable in its electronic form.
Please observe the information given on www.metas.ch/ecert

OIML Certificate No R076/2006-A-CH1-2019.06-01

Metrological characteristics

ME..T instruments

Accuracy class	Ⓘ	Ⓜ
Max	≤ 220 g	≤ 5 200 g
d	≥ 0.1 mg	≥ 1 mg
e	≥ 1 mg	≥ 10 mg
n	≤ 220 000	≤ 52 000
Tare-weighing range	Until 100 % of Max	
Temperature range	10 °C / 30 °C	

JET instruments

Accuracy class	Ⓘ	Ⓜ
Max	≤ 12 200 g	≤ 6 200 g
d	≥ 0.1 mg	≥ 1 mg
e	≥ 1mg	≥ 10 mg
n	≤ 160 000	≤ 62 000
Tare-weighing range	Until 100 % of Max	
Temperature range	10 °C / 30 °C	

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
6030-00039	11.07.2016	38
6030-00040	11.07.2016	49
6030-00041	11.07.2016	49
6030-00044	12.07.2016	23
133-08543	03.09.2019	57
133-08544	03.09.2019	57
6030-01228-01	26.01.2021	11

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
R076_2006-A-CH1-2019_06_CH_MT_ME-T-JET_Doc_List	27.09.2019	4

OIML Certificate No R076/2006-A-CH1-2019.06-01

OIML Certificate History:

Revision No.	Date	Description of the modification
00	03.09.2019	First issue, revision of R76/2006-CH1-16.01
01	26.01.2021	Editorial change: including the value of the scale interval d

The OIML Issuing Authority CH1

3003 Bern-Wabern, 26 January 2021

Approved by

Gulian Couvreur, Head of sector
METAS-Cert



Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full