



Member State
Switzerland

OIML Certificate No.
R076/2006-A-CH1-2021.01

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

Name **Federal Institute of Metrology METAS**
Conformity Evaluation Body METAS-Cert

Address Lindenweg 50, 3003 Bern-Wabern, Switzerland

Person responsible Gulian Couvreur, Head of METAS-Cert

Applicant

Name **Haenni Instruments Inc.**

Address Industrie Neuhof 66, 3422 Kirchberg, Switzerland

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified type

Indicator / Processing Unit

Type **EC 120**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R 76-1, edition 2006

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Annex to the OIML Type Evaluation Report No R076/2006-A-CH1-2021.01

Metrological characteristics

Temperature range -10 °C / 50 °C

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
6030-01652	2021-04-27	8

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
R076/2006-A-CH1-2021.01_LERD	2021-04-27	3

OIML Certificate History:

Revision No.	Date	Description of the modification
00	2021-04-27	First issue

The OIML Issuing Authority CH1

3003 Bern-Wabern, 2021-04-27

Approved by

Gulian Couvreur, Head of sector
METAS-Cert



Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full