

Member State  
P.R.China



OIML Certificate No.  
R76/2006-CN1-11.01

# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 P.R. China

## Applicant

Name: Shang Hai Digital Balance Electronic Co., LTD

Address: #788 Song Xiu RD Qing Pu Districal of Shanghai (201703) P.R.China

## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

## Identification of the certified type

Price Labelling Scale

| Type                    | BCS-100PEA      |       |
|-------------------------|-----------------|-------|
| Accuracy class          |                 |       |
| Max                     | 15 kg           | 30 kg |
| Min                     | 100 g           | 200 g |
| e                       | 5 g             | 10 g  |
| n                       | 3000            |       |
| Temperature range(°C)   | 0 °C ~ 40 °C    |       |
| Power supply voltage    | AC 220 V, 50 Hz |       |
| Software identification | 20100831        |       |

The numbers of the associated test reports and the included pages:

NIM-LSmc-1101      55 pages pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 2006

For accuracy class (if applicable): ④

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority  
Office of OIML Affairs



October 30, 2011

The CIML member

October 30, 2011

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.