

Member State
P.R.China



OIML Certificate No.
R76/2006-CNI-13.02

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 P.R.China

Applicant

Name: Guangdong Huapu Electrical Appliance Group Co., Ltd
Address: Lunjiao Section, #105 National Highway, Shunde,
Guangdong, (528308) P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Weighing Indicator

Type:	XK3119, XK3119-WM(B)
Accuracy class:	III
Maximum number of verification scale intervals n:	6000
Operating temperature range(°C):	0 °C ~ 40 °C
Value of the fractional error ₀ :	0,5
Power supply voltage:	AC 100 V ~ 110 V, AC 220 V ~ 240 V/ 50 Hz DC 6V
Load cell excitation voltage:	5 V
Maximum signal voltage for dead load (mV):	15 mV
Minimum input-voltage per verification scale interval e (µV):	1 µV
Measuring range maximum voltage (mV):	15 mV
Minimum load cell impedance(Ω):	87Ω
Maximum load cell impedance(Ω):	1000 Ω

The numbers of the associated test reports and the included pages:

NIM-LSmc-1203 50pages, NIM-LSmc-1204 39pages

OIML Certificate No.
R76/2006-CN1-13.02

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 2006

For accuracy class (if applicable): Ⅲ

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority
Office of OIML Affairs



March 3, 2013

The CIML member

蔺长斌

March 3, 2013