

Member State  
P.R.China



OIML Certificate No.  
R76/2006-CN1-13.05  
Revision 1

# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

Address: No.9 Madiandonglu, Haidian District, Beijing, 100088, P. R. China

## Applicant

Name: Shanghai Yaohua Weighing System Co., LTD

Address: No. 4059, Shangnan Rd. Pudong District, Shanghai (200124) P. R. China

## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

## Identification of the certified type

### Weighing Indicator

Type	YH-T7	YH-T7+E
Accuracy class	III	
Maximum number of verification scale intervals (n)	3000	
Value of the fractional error ( p <sub>i</sub> )	0.5	
Minimum input-voltage per verification scale interval e ( μV )	1 μV	
Minimum load cell impedance(Ω)	345 Ω	
Maximum load cell impedance(Ω)	700 Ω	
Operating temperature range(°C)	0 °C ~ 40 °C	

The numbers of the associated test reports and the included pages:

NIM-LSmc-1303 38 pages, NIM-LSmc-1304 49 pages

This certificate attests the conformity of the above-mentioned type (represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 2006

For accuracy class (if applicable): Ⅲ

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

This certificate replaces the previous versions and the reason for the revision is typing mistake of the type of instrument.

The issuing authority

Office of OIML Affairs

中国秘书处

January 15, 2014

The OIML member

蒲长城

January 15, 2014

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.