

Member State  
P.R.China

OIML Certificate No.  
R76/2006-CN1-16.02  
Revision 1



# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine  
Address: No.9 Madiandonglu, Haidian District, Beijing, 100088 P.R.China

## Applicant

Name: Mettler Toledo (Changzhou) Measurement Technology Ltd.  
Address: No. 111, west Taihu Road, Xinbei District, Changzhou, Jiangsu. (213125) P.R.China

## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

## Identification of the certified type

Terminal, as part of a non-automatic weighing instrument

Type	ICS425	ICS435	ICS445	ICS465	ICS685
Accuracy class	Ⓜ				
Maximum number of verification scale interval $n$	100000				
Fraction of the maximum permissible error ( $p_i$ )	0				
weighing range	Single interval / Multi-interval / Multi range				
Power supply	AC 100-240 V, 50/60 Hz				
Software identification	XY-XX-01.YY.YY-XX-YY (X is a character from A to Z; Y is a number between 0 to 9.)				
List of devices:	semi-automatic zero-setting device, automatic zero-setting device, initial zero-setting device, tare balancing or tare-weighing device.				

The numbers of the associated test reports is NIM-LSmc-1604 including 15 pages, NIM-LSmc-1605 including 15 pages and NIM-LSmc-1606 including 29 pages.

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 2006

For accuracy class (if applicable): ②

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports. This certificate replaces the previous version and the reason for a typing mistake.

The issuing authority

Office of OIML Affairs

中国秘书处

March 20, 2017

The CIML member

吴清江

March 20, 2017