

Member State
P.R.China

OIML Certificate No.
R76/2006-CN1-16.03



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: No. 9 Madiandonglu, Haidian District, Beijing, 100088 P.R. China

Applicant

Name: CHONGQING DATA CONTROL TECHNOLOGY CO., LTD.
Address: No. 2, Fengxi Road, Caijiagang Town, Beibei District, Chongqing (400707) P.R. China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Electronic truck scale

| Type | SCS-100t | SCS-150t |
|--|---|----------|
| Accuracy class | III | |
| Max | 100 t | 150 t |
| Min | 1 t | |
| Verification scale interval <i>e</i> | 50 kg | |
| Number of verification scale interval <i>n</i> | 2000 | 3000 |
| Temperature range | +0 °C ~ +40 °C | |
| Humidity range | ≤85 %RH | |
| Software identification | A 17-C | |
| List of devices: | Zero-setting device: Semi-automatic, Automatic zero-setting, Initial zero-setting, Zero-tracking; Tare-device: Tare balancing, Subtractive tare. | |

The numbers of the associated test reports is NIM-LSmc-1607 including 14 pages, NIM-LSmc-1608 including 27 pages .

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 2006

For accuracy class (if applicable): Ⅲ

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority

Office of OIML Affairs

中国秘书处

November 15, 2016

The CIML member

吴清江

November 15, 2016

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.