

Member State
P.R.China

OIML Certificate No.
R76/2006-CN1-17.01



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

Address: No.9 Madiandonglu, Haidian District, Beijing, 100088 P.R.China

Applicant

Name: Mettler Toledo (Changzhou) Measurement Technology Ltd.

Address: No. 111, west Taihu Road, Xinbei District, Changzhou, Jiangsu. (213125) P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type : Forklift scale

Type	VFS120 series include VFS120-1.5- II -762XX, VFS120-1.5- II -920XX, VFS120-1.5- II -1000XX, VFS120-1.5- II -1040XX, VFS120-2- II -920XX, VFS120-2- II -1000XX, VFS120-2- II -1040XX, VFS120-2- II -1150XX, VFS120-3- III -1100XX, VFS120-3- III -1150XX, VFS120-5- III -1250XX, VFS120-5- IV -1250XX (XX represent the sales area code)
Accuracy class	III
Max	1500 kg~5000 kg
Min	20 kg~100 kg
Verification scale interval e	1 kg~5 kg
Number of verification scale interval n	1000~1500
Temperature range	-10 °C~40 °C
Humidity range	≤85 %RH
Software identification	1.XX.00YY (XX and YY are a number between 00 to 99.)
List of devices:	Zero-setting device: Semi-automatic, Initial zero-setting, Zero-tracking; Tare-device: Tare balancing, Subtractive tare.

The numbers of the associated test reports and the included pages:

NIM-LSmc-1502 including 37 pages and NIM-LSmc-1503 including 26 pages, NIM-LSmc-1609 including 26 pages and NIM-LSmc-1610 including 15 pages.


The compatibility checking related OIML test reports of load cell are as below:

LSfc2014-6005, that includes 18 pages, LSfc2014-6006, that includes 16 pages.

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 2006

For accuracy class (if applicable): 

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority


Office of OIML Affairs

中国秘书处

September 8, 2017

The OIML member



September 8, 2017

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.