



OIML Member State
P. R. China



OIML Certificate No.
R76/2006-A-CN2-2018.01

OIML CERTIFICATE

ISSUED UNDER SCHEME A

OIML Issuing Authority Name: National Institute of Metrology, China
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist, Beijing, P.R.China
Person responsible: Mr. Fang Xiang, Director

Applicant Name: Mettler Toledo (ChangZhou) Measurement Technology Ltd.
Address: No. 111, West Taihu Road, Xinbei District, Changzhou, Jiangsu, P.R. China.

Manufacturer Name: Mettler Toledo (ChangZhou) Measurement Technology Ltd.
Address: No. 111, West Taihu Road, Xinbei District, Changzhou, Jiangsu, P.R. China.

Identification of the certified type Terminal for Non-Automatic Weighing Instrument
IND 780 series (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class: **II**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 14 August 2018

Fang Xiang
Director
National Institute of Metrology, China



No.18, Bei San Huan Dong Lu,
Chaoyang Dist, Beijing,
100029, P.R.China
+86 10 64524602
oimlia@nim.ac.cn
www.nim.ac.cn

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R76/2006-CN2-2018.01 dated 7 August 2018 that includes 24 pages

and the associated OIML test report:

No. LSmt2017-1582-1 dated 5 August 2018 that includes 17 pages

No. LSmt2017-1582-2 dated 5 August 2018 that includes 17 pages

OIML Certificate History

Revision No.	Date	Description of the modification
0	14 August 2018	Initial issue

Characteristics of the Non-automatic Weighing Instrument

Type	IND 780 Harsh and IND 780 Panel
Accuracy class	Ⓜ
Tare range	-Max
Maximum number of verification scale interval n	100000
Fraction of the maximum permissible error (p_i)	0.0
Power supply	AC (100-240) V, 50/60 Hz
Software identification	8.X.YY (8 is related to the legal metrological relevant software, and X and Y are numbers between 0 to 9 related to the non- legal metrological relevant software.)