



OIML Member State
P. R. China



OIML Certificate No.
R76/2006-A-CN2-2020.01

OIML CERTIFICATE

ISSUED UNDER SCHEME A

OIML Issuing Authority Name: National Institute of Metrology, China
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist, Beijing, P.R.China
Person responsible: Mr. Fang Xiang, Director

Applicant Name: SHENZHEN GENERAL MEASURE TECHNOLOGY CO., LTD.
Address: A1-3/F, Block C, Tongfang Information Harbor, No.11 Langshan Rd,
Nanshan District, Shenzhen, P.R. China

Manufacturer Name: ZHUZHOU GENERAL MEASURE TECHNOLOGY CO., LTD.
Address: 1st-3rd Floor, Block #B1, NO.959 Tianyi Avenue, Hi-Tech Xinma Golden
Valley, Tianyuan District, Zhuzhou, Hunan Province, P.R. China

Identification of the certified type Weighing Indicator GMT-P1 (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 – Edition 2006

For accuracy class:



This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 4 November 2020



No.18, Bei San Huan Dong Lu,
Chaoyang Dist, Beijing,
100029, P.R.China
+86 10 64525646
oimlia@nim.ac.cn
www.nim.ac.cn

Fang Xiang
Director
National Institute of Metrology, China

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R76/2006-CN2-2020.01 dated 2 November 2020 that includes 25 pages

and the associated OIML test report:

No. LSmm2020-04329 dated 24 August 2020 that includes 42 pages

OIML Certificate History

Revision No.	Date	Description of the modification
0	4 November 2020	Initial issue

Characteristics of the Instrument

Type	GMT-P1
Accuracy class	III
Value of the error fraction p_i	0.5
Minimum input-voltage per verification scale interval, e	1 μ V
Maximum number of verification scale intervals, n	6000
Temperature range	-10 °C ~ +40 °C
Power supply	DC 24 V
Load cell excitation voltage	5 V
Minimum signal voltage for dead load	0.2 mV
Maximum signal voltage for dead load	8 mV
Measuring range minimum voltage	0.2 mV
Measuring range maximum voltage	15 mV
Minimum load cell impedance	350 Ω
Maximum load cell impedance	1000 Ω
Sense system	with (six wire)
Software version	01.01.00
Maximum cable length	2144.83 m/mm ²