

OIML Member State P. R. China



OIML Certificate No. R76/2006-A-CN2-21.02

OIML CERTIFICATE

ISSUED UNDER SCHEME A

OIML Issuing Authority Name:

National Institute of Metrology, China

Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China

Person responsible: Mr. Fang Xiang, Director

Applicant

Name: XIAMEN HANIN ELECTRONIC TECHNOLOGY CO.,LTD

Address: Room 305A, Angye Building, Pioneering Park, Torch High-tech Xiamen,

Fujian, China

Manufacturer

Name: XIAMEN HANIN ELECTRONIC TECHNOLOGY CO.,LTD

Address: Room 305A, Angye Building, Pioneering Park, Torch High-tech Xiamen,

Fujian, China

Identification of the

Electronic weighing scale C-S1 (Further characteristics on page 2)

certified type

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006

For accuracy class:



This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority:

National Institute of Metrology, China

Date:

08 November 2021

Director

Fang Xiang

National Institute of Metrology, China

No.18, Bei San Huan Dong Lu, Chaoyang Dist, Beijing, 100029, P.R.China +86 10 64525646 oimlia@nim.ac.cn www.nim.ac.cn

OIML Member State P. R. China

OIML Certificate No. R76/2006-A-CN2-21.02

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R76/2006-CN2-21.02

dated

dated

04 November 2021 that includes

pages

and the associated OIML test report:

No. LSmm2021-13749

18 October 2021

that includes

pages

OIML Certificate History

Revision No.	Date	Description of the modification
0	08 November 2021	Initial issue

Characteristics of the Instrument

Type	C-S1
Accuracy class	
Max	6kg/15kg
Min	40g
e	2g/5g
d	2g/5g
n	3000/3000(multi-interval)
Tare-balancing range	-5.998kg
Temperature range	-10°C ~ +40 °C
Power supply	AC 220V, 50Hz
Software version	V 1.008
Catio	n 515