



OIML Member State  
P. R. China



OIML Certificate No.  
R76/2006-A-CN2-21.03

## OIML CERTIFICATE

### ISSUED UNDER SCHEME A

**OIML Issuing Authority** Name: National Institute of Metrology, China  
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China  
Person responsible: Mr. Fang Xiang, Director

**Applicant** Name: Zhongshan Camry Electronic Co., LTD  
Address: Baishawan Industrial Park, Qiwan Road East, East District, Zhongshan, Guangdong, China

**Manufacturer** Name: Zhongshan Camry Electronic Co., LTD  
Address: Baishawan Industrial Park, Qiwan Road East, East District, Zhongshan, Guangdong, China

**Identification of the certified type** Price Computing Scale ACS-15, ACS-30 (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76 – Edition 2006**

For accuracy class: III

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 31 December 2021

**Fang Xiang**  
Director  
National Institute of Metrology, China



No.18, Bei San Huan Dong Lu,  
Chaoyang Dist, Beijing,  
100029, P.R.China  
+86 10 64525646  
oimlia@nim.ac.cn  
www.nim.ac.cn

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R76/2006-CN2-21.03 dated 31 December 2021 that includes 24 pages  
and the associated OIML test report:

No. LSmm2021-14439 dated 07 December 2021 that includes 43 pages

### OIML Certificate History

Revision No.	Date	Description of the modification
0	31 December 2021	Initial issue

### Characteristics of the Instrument

Type	ACS-15	ACS-30
Accuracy class	III	
Max	15kg	30 kg
Min	100g	200g
<i>e</i>	5g	10g
<i>d</i>	5g	10g
<i>n</i>	3000	
Tare-balancing range	-15 kg	-30 kg
Temperature range	0 °C ~ +40 °C	
Power supply	AC 100 ~240 V, 50/60 Hz, DC/rechargeable battery 4 V	
Software version	Vr 1.68	