



OIML Member State
P. R. China



OIML Certificate No.
R76/2006-A-CN2-22.01

OIML CERTIFICATE

ISSUED UNDER SCHEME A

OIML Issuing Authority Name: National Institute of Metrology, China
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China
Person responsible: Mr. Fang Xiang, Director

Applicant Name: XIAMEN HANIN ELECTRONIC TECHNOLOGY CO.,LTD
Address: Room 305A,Angye Building, Pioneering Park, Torch High-tech Xiamen, Fujian, China

Manufacturer Name: XIAMEN HANIN ELECTRONIC TECHNOLOGY CO.,LTD
Address: Room 305A,Angye Building, Pioneering Park, Torch High-tech Xiamen, Fujian, China

Identification of the certified type Bar code label scale C-L1 Series (C-L106、C-L115、C-L130) (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 – Edition 2006

For accuracy class: III

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 09 June 2022



No.18, Bei San Huan Dong Lu,
Chaoyang Dist, Beijing,
100029, P.R.China
+86 10 64525646
oimlia@nim.ac.cn
www.nim.ac.cn

Fang Xiang
Director
National Institute of Metrology, China

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R76/2006-CN2-22.01 dated 08 June 2022 that includes 25 pages
and the associated OIML test report:

No. LSmm2022-00922 dated 12 April 2022 that includes 43 pages

OIML Certificate History

Revision No.	Date	Description of the modification
0	09 June 2022	Initial issue

Characteristics of the Instrument

Type	C-L1		
Variants	C-L106	C-L115	C-L130
Accuracy class	III		
Max	3 kg/6 kg	6 kg/15 kg	15 kg/30 kg
Min	20 g	40 g	100 g
<i>e</i>	1 g/2 g	2 g/5 g	5 g/10 g
<i>d</i>	1 g/2 g	2 g/5 g	5 g/10 g
<i>n</i>	3000/3000 (multi-interval)		
Tare-balancing range	-2.999 kg	-5.998	14.995
Temperature range	0 °C ~ +40 °C		
Power supply	AC 220 V, 50 Hz		
Software version	V 1.105		