

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate N°
R76/2006-DE1-10.01

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Dirk Ratschko

Applicant

Name: Bizerba GmbH & Co. KG
Address: Wilhelm-Kraut-Str. 65, 72336 Balingen

Manufacturer of the certified type is the applicant.

Identification of the certified type Non-automatic weighing instrument for direct sales to the public
Type: SC II...

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006
for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R76/2006-DE1-10.01

The conformity was established by the results of tests and examinations provided in the Report No. 1.12-4044914 and the associated Test Reports

No. 1.12-4044914/1 that includes 54 pages
No. 1.12-4044914/2 that includes 55 pages
No. 1.12-4044914/3 that includes 55 pages

The Issuing Authority



Dr. D. Ratschko
Oberregierungsrat

2010-06-09

The CIML Member



Dr. R. Schwartz
Direktor und Professor

2010-06-09

L.S.

Table 1

Accuracy class	III
Type	SC II...
Max \leq	300kg
e =	1 g ... 50 g
d =	1 g ... 50 g
n \leq	7500
Tare balancing range (subtractive)	≤ 100 % of Max
Preset tare range	\leq Max \leq Max ₁ at multi-interval instruments
Temperature range	-10 °C / 40 °C

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.