

Member State of OIML
Germany



OIML Certificate No.
R76/2006-DE1-10.03

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Dirk Ratschko

Applicant

Name: Liebherr-Mischtechnik GmbH
Address: Im Elchgrund 12
88427 Bad Schussenried
Germany

Manufacturer of the certified type is the applicant.

Identification of the certified type Non-automatic electromechanical weighing instrument with or without lever system

Type: Liebherr Litronic-MPS III SWS 3

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006

for accuracy classes  and 

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

OIML Certificate No.
R76/2006-DE1-10.03

The conformity was established by tests described in the report N° 1.12-4043670 and the associated test report No. 1.12-4043670/1 (45 pages).

The Issuing Authority



Dr. D. Ratschko
Head of Department

08.10.2010



The OIML Member



Dr. R. Schwartz
Head of Division

08.10.2010

Table 1

Accuracy class	(III)	(III)
Max	1 kg ... 50 t	1 kg ... 100 t
Number of verification scale intervals	≤ 6000	≤ 1000
Temperature range	-10 °C bis +40 °C	

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.