

Member State of OIML  
Germany



OIML Certificate No.  
**R76/2006-DE1-12.02**  
Revision 1

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. O. Mack

### Applicant

Name: Bizerba GmbH & Co. KG  
Address: Wilhelm-Kraut-Str. 65,  
72336 Balingen  
Germany

### Manufacturer

Name:	Bizerba GmbH & Co. KG	Bizerba (Shanghai) WeighTech & Systems Co., Ltd.
Address:	Wilhelm-Kraut-Str. 65 72336 Balingen Germany	D-3, No. 195 Qianpu Road, East New Area of Songjiang Industrial Zone, Shanghai, 201611, P.R. China

### Identification of the certified type

Non-automatic electromechanical price-computing weighing instrument  
for direct sales to the public  
Type: XC...

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76-1**, edition 2006  
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

OIML Certificate No.  
**R76/2006-DE1-12.02**  
**Revision 1**

This Revision 1 was issued because new variants of load-receptors, software updates and structural modifications were added. The conformity is established in the Report No. 1.12-4056935, Revision 1 (21 pages) and by tests described in the associated Test Reports

- No. 1304061104 that includes 16 pages
- No. 1304071104 that includes 16 pages
- No. 1403012104 that includes 56 pages
- No. 1403021104 that includes 16 pages
- No. 1006021104 that includes 43 pages
- No. 1111161104 that includes 56 pages
- No. 1201021104 that includes 12 pages
- No. 1204031104 that includes 55 pages
- No. 1409011104 that includes 51 pages

The test results of the former Test Reports No. 1.12-4056935/1 and 1.12-4056935/2 remain valid.

Technical Data:

Type	XC...
Accuracy class	III
Max	≤ 300 kg
e =	1 g ... 50 g
n ≤	7500
Tare-balancing range ≤	100 % of Max
Temperature range	- 10 °C / + 40 °C

**The Issuing Authority**



Dr. O. Mack  
 Member of Certification Body

26.02.2016



**The OIML Member**



Dr. R. Schwartz  
 Vice President

26.02.2016

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.