



**Physikalisch-Technische Bundesanstalt**  
**Braunschweig und Berlin**  
Nationales Metrologieinstitut

Member State of OIML  
Germany



OIML Certificate No.  
**R76/2006-DE1-12.02**  
**Revision 2**

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. O. Mack

### Applicant

Name: Bizerba SE & Co. KG  
Address: Wilhelm-Kraut-Str. 65,  
72336 Balingen  
Germany

### Manufacturer

Name: Bizerba SE & Co. KG  
Address: Wilhelm-Kraut-Str. 65  
72336 Balingen  
Germany  
Bizerba (Shanghai) WeighTech & Systems Co., Ltd.  
D-3, No. 195 Qianpu Road,  
East New Area of Songjiang Industrial Zone,  
Shanghai, 201611,  
P.R. China

### Identification of the certified type

Non-automatic electromechanical price-computing weighing instrument  
for direct sales to the public  
Type: XC...

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76-1**, edition 2006  
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



**OIML Certificate No.**  
**R76/2006-DE1-12.02**  
**Revision 2**

The Revision 1 was issued because new variants of load-receptors, software updates and structural modifications were added.

This Revision 2 was issued because a new type and new variants of weighing modules were added.

New test reports were added:

- No. 1104011104 that includes 55 pages
- No. 1104081104 that includes 16 pages
- No. 1405011104 that includes 16 pages
- No. 1509011104 that includes 56 pages

The conformity is established in the Report No. 1.12-4056935, Revision 2 (25 pages) and by tests described in the associated Test Reports

The test results of the former Test Reports No. 1.12-4056935/1, No. 1.12-4056935/2, No. 1304061104, No. 1304071104, No. 1403012104, No. 1403021104, No. 1006021104, No. 1111161104, No. 1201021104, No. 1204031104 and No. 1409011104 remain valid.

Technical Data:

|                        |                   |
|------------------------|-------------------|
| Type                   | XC...             |
| Accuracy class         | III               |
| Max                    | ≤ 300 kg          |
| e =                    | 1 g ... 50 g      |
| n ≤                    | 7500              |
| Tare-balancing range ≤ | 100 % of Max      |
| Temperature range      | - 10 °C / + 40 °C |

**The Issuing Authority**

Dr. O. Mack  
 Member of Certification Body

04.04.2017



**The OIML Member**

Dr. R. Schwartz  
 Vice President

04.04.2017

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.