

Member State of OIML
Germany



OIML Certificate No.
R76/2006-DE1-12.02

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Oliver Mack

Applicant

Name: Bizerba GmbH & Co. KG
Address: Wilhelm-Kraut-Str. 65,
72336 Balingen
Germany

Manufacturer

| | | |
|----------|--|---|
| Name: | Bizerba GmbH & Co. KG | Bizerba (Shanghai) WeighTech & Systems Co., Ltd. |
| Address: | Wilhelm-Kraut-Str. 65 72336 Balingen Germany | D-3, No. 195 Qianpu Road, East New Area of Songjiang Industrial Zone, Shanghai, 201611, P.R. China |

Identification of the certified type Non-automatic, electromechanical price-computing weighing instrument for direct sales to the public
Type: XC...

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006,
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

OIML Certificate No.
R76/2006-DE1-12.02

The conformity was established by the Report 1.12-4056935 and the results of tests and examinations provided in the associated Test Reports

No. 1.12-4056935/1 that includes 54 pages and
No. 1.12-4056935/2 that includes 54 pages

Technical Data:

| | |
|------------------------|-------------------|
| Type | XC... |
| Accuracy class | III |
| Max | ≤ 30kg |
| e = | 1 g ... 10 g |
| n ≤ | 6000 |
| Tare-balancing range ≤ | 100 % of Max |
| Temperature range | - 10 °C / + 40 °C |

The Issuing Authority



M. Denzel
Member of Certification Body

18.12.2012



The OIML Member



Dr. R. Schwartz
Head of Division

18.12.2012

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.