

Member State of OIML  
Germany



OIML Certificate No.  
R76/2006-DE1-12.04

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. Oliver Mack

### Applicant

Name: OAS AG  
Address: Linzer Str. 7, 28359 Bremen

Manufacturer of the certified type is the applicant.

**Identification of the certified type** Non automatic electromechanical weighing instrument

Type: Wcon-I

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006,  
for accuracy classes  and 

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

OIML Certificate No.  
R76/2006-DE1-12.04

The conformity was established by the Report 1.12-4051047 and the results of tests and examinations provided in the associated Test Reports

No. 1.12-4051046/1                    that includes 46 pages  
 No. 1.12-4051046/2                    that includes 20 pages  
 No. 1.12-4051046/3                    that includes 20 pages

Technical data:

Accuracy class	III	III
Maximum capacity Max	1 kg ... 200 t	
Number of verification scale intervals <sup>1)</sup>	≤ 6000 <sup>2)</sup>	≤ 1000
Tare-balancing range (subtractive)	≤ 100 % • Max	
Tare-balancing range (additive)	≤ 100 % • Max	

<sup>1)</sup> This applies to each range of single- and multiple range instruments.

<sup>2)</sup> Using a lever system, the maximum number of verification scale intervals is limited to  $n \leq 3000$ .

**The Issuing Authority**



M. Denzel  
Member of Certification Body

17.12.2012



**The OIML Member**



Dr. R. Schwartz  
Head of Division

17.12.2012

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.