

Member State of OIML  
Germany



OIML Certificate No.  
R76/2006-DE1-13.01

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. Oliver Mack

### Applicant

Name: Schenck Process GmbH  
Address: Pallaswiesenstr. 100, 64293 Darmstadt

Manufacturer of the certified type is the applicant.

**Identification of the certified type**    Indicating and operating terminal  
Type: VOP280xx

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76-1**, edition 2006

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

OIML Certificate No.  
**R76/2006-DE1-13.01**

The conformity was established by the Report 1.12-4057200 and the results of tests and examinations provided in the associated Test Report

No. 1.12-4057200/1 that includes 9 pages

### TECHNICAL DATA

Indicating and operating terminal , type VOP280xx

Kind of connection for weighing instruments		Ethernet, RS 485 or RS 232
Limits of temperature range	$T_{min} \dots T_{max}$	-10°C ... +40°C
Power supply VOP28000 and VOP28010 VOP28030	$U_p$	10 - 36 V DC, 85 - 264 V AC

#### The Issuing Authority




Dr. Oliver Mack  
Head of Working Group

27.02.2013



#### The OIML Member



Dr. R. Schwartz  
Head of Division

27.02.2013

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.