

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate No.
R76/2006-DE1-13.02

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: M. Denzel

Applicant

Name: seca gmbh & co. kg.
Address: Hammer Steindamm 9-25,
22089 Hamburg
Germany

Manufacturer of the certified type is the applicant.

Identification of the certified type Non-automatic electromechanical baby weighing instrument
Type: BIS03A

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006,
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate No.
R76/2006-DE1-13.02

The conformity was established by the results of tests and examinations provided in the associated Test Reports

No. 1.12-4061179/1 that includes 56 pages

The Issuing Authority

The OIML Member

M. Denzel
Member of Certification Body

Dr. R. Schwartz
Head of Division

27.05.2013

27.05.2013

Identification of the pattern (continued)

Non-automatic electromechanical weighing instrument without lever system. The weighing range with Max, Min, e, d and number of verification scale intervals is shown in table 1:

Table 1: Technical data

Type	BIS03A
Accuracy class	III
Min	0.2 kg
Max	15 kg
e=d	0.01 kg
n	1500
Tare balancing range, subtractive	10 kg
Temperature range	10 °C ... 40 °C

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.