



Member State of OIML
Germany



OIML Certificate No.
R76/2006-DE1-15.04

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. O. Mack

Applicant

Name: Siemens AG
Address: Östliche Rheinbrückenstr.-50, 76187 Karlsruhe

Manufacturer of the certified type is the applicant.

Identification of the certified type

Nichtselbsttätige elektromechanische Waage mit oder ohne Hebelwerk
Non-automatic electromechanical weighing instrument with or without
lever
Type: SIWAREX WP231 NAWI

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006

for accuracy class(es)  

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports

No. 1.12-4073499/1 that includes 39 pages
 No. 1.12-4073499/2 that includes 14 pages
 No. 1.12-4073499/3 that includes 8 pages

The Issuing Authority



M. Denzel
 Member of Certification Body

11.12.2015



The OIML Member



Dr. R. Schwartz
 Vice President

11.12.2015

Technical data:

Accuracy class	III	III
Maximum capacity Max	0,1 kg ... 400 t	4 kg ... 400 t
Number n of scale intervals	≤ 3000	≤ 1000
Tare-balancing range (subtractive)	≤ 100 % • Max	
Tare-balancing range (additive)	≤ 250 % • Max	
Preset tare range (subtractive)	≤ 100 % • Max	
Preset tare range (additive)	≤ 250 % • Max	
Temperature range	-10 °C ... +40 °C	

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.