



Physikalisch-Technische Bundesanstalt
Braunschweig und Berlin
Nationales Metrologieinstitut

Member State of OIML
Germany



OIML Certificate No.
R76/2006-DE1-16.01
Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dipl.-Ing. (FH) T. Schwabe

Applicant

Name: Bizerba SE & Co. KG
Address: Wilhelm-Kraut-Str. 65, 72336 Balingen

Manufacturer of the certified type is:

Bizerba SE & Co. KG
Wilhelm-Kraut-Str. 65
72336 Balingen

Identification of the certified type

Non-automatic price-computing weighing instrument for direct sales to the public
Type: MC...

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006
for accuracy class **III**



OIML Certificate No.
R76/2006-DE1-16.01
Revision 1

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports

No. 1.12-4073750/1 (1 st revision)	that includes 13 pages
No. 1509031104	that includes 17 pages
No. 1504012104	that includes 56 pages

The Issuing Authority

Dipl.-Ing. (FH) T. Schwabe
 Member of Certification Body

26.10.2017



The OIML Member

Hon.-Prof. Dr. R. Schwartz
 Vice President

26.10.2017

Identification of the pattern (continued)

Weighing instrument with load receptor, terminal with touch, keyboard and intern or extern printer.
 The weighing ranges with Max, Min, e and number of verification scale intervals may be chosen within the limits of No. 3.2 of R76-1 and of Table 1.

Table 1

Accuracy class	III
Maximum capacity Max	30 kg
Verification scale interval e	1 g ... 10 g
Number of verification scale intervals	≤ 6000
Tare-balancing range (subtractive)	≤ 100 % • Max
Preset tare range	≤ 100 % • Max
Temperature range	-10 °C / +40 °C

OIML Certificate No.
R76/2006-DE1-16.01
Revision 1

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.