



Physikalisch-Technische Bundesanstalt
Braunschweig und Berlin

OIML Member State
Germany

OIML Certificate No.
R76/2006-A-DE1-2019.06

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority

Name: Physikalisch-Technische Bundesanstalt,
Conformity Assessment Body
Address: Bundesallee 100, 38116 Braunschweig, GERMANY
Person responsible: Hon.-Prof. Dr. R. Schwartz

Applicant

Name: Erte Endüstriyel Elektronik San.ve TIC.LTD.STI
Address: 1443 Sok.Tesisat Is Merkezi, No:2R, 35170 Izmir Turkey

Manufacturer

Name: Erte Endüstriyel Elektronik San.ve TIC.LTD.STI
Address: 1443 Sok.Tesisat Is Merkezi, No:2R, 35170 Izmir Turkey

Identification of the certified type *(the detailed characteristics will be defined in the additional pages)*

Non-automatic electromechanical weighing instrument with or without lever system
Type: B3

Designation of the module *(if applicable)*

Not applicable

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76

Edition (year): 2006

For accuracy class (if applicable): III, IIII

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. 1.12-4092401 dated 02.07.2019 that includes 19 pages

The technical documentation relating to the identified type is contained in documentation file:

No. ZDS-R76/2006-A-DE1-2019.06 dated 02.07.2019 that includes 3 pages

OIML Certificate History

Revision No.	Date	Description of the modification
0	02.07.2019	Initial issuing

Identification, signature and stamp

The Issuing Authority

The CIML Member

Dr. Oliver Mack

Member of Conformity Assessment Body



Hon.-Prof. Dr. R. Schwartz

Vice President of PTB

Date: 02.07.2019

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Identification of the certified type (continued)

Metrological characteristics of the pattern:

Accuracy class		(III)	(III)
Minimum load Min		$\geq 20 e$	
Maximum capacity Max	t	≤ 300	
Verification scale interval e	g	≥ 0.1	
Number of verification scale intervals n		≤ 10000	≤ 1000
Number of verification scale intervals ni		$\leq 2 \times 6000$	$\leq 2 \times 1000$
Tare range (subtractive)	• Max	$\leq 100\%$	
Preset tare range	• Max	$\leq 100\%$	