



**OIML Member State**  
Denmark

**OIML Certificate No.**  
**R76/2006-A-DK2-2020.06**

**OIML CERTIFICATE ISSUED UNDER SCHEME A**

**OIML Issuing Authority**

Name: **FORCE Certification A/S**  
Address: **Park Allé 345, 2605 Brøndby, Denmark**  
Person responsible: **Leif Madsen**

**Applicant**

Name: **CAS Corporation**  
Address: **#262, Geurugogae-ro, Gwangjeok-myeon,  
Yangju-si, Gyeonggi-do  
REPUBLIC OF KOREA**

**Manufacturer**

**CAS (Zhejiang) Electronics Co. Ltd, China.**  
**CAS Corporation, Republic of Korea**  
**CAS Elektronik San. Tic. A.S., Turkey**  
**CAS Deutschland AG, Germany.**

**Identification of the certified type** (*the detailed characteristics will be defined in the additional pages*)

**PBII**

**Designation of the module** (*if applicable*)

**Non-automatic electronic weighing instrument**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1, Edition (year): 2006**

For accuracy class (if applicable): **III**

**OIML Certificate No.  
R76/2006-A-DK2-2020.06**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type examination report: No. 120-22358.10, dated 27 April 2020, that includes 69 pages

Type evaluation report: No. 120-22358.90.20 dated 06 May 2020, that includes 20 pages

The technical documentation relating to the identified type is contained in documentation file:  
120-22358

**OIML Certificate History**

| <b>Revision No.</b> | <b>Date</b>  | <b>Description of the modification</b> |
|---------------------|--------------|--|
| Initial version     | 02 June 2020 | -                                      |
|                     |              |  |
|                     |              |  |
|                     |              |  |

Identification, signature and stamp

**The OIML Issuing Authority**

FORCE Certification A/S

Date: 02 June 2020

Jens Hovgård Jensen  
Certification Manager

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

## Descriptive annex

### Characteristics

- Accuracy class III
- Single interval, multi interval (dual)
- Maximum number of verification scale intervals: 3000 or 2x3000
- Maximum capacity (Max): 30 kg to 150 kg
- Minimum capacity (Min):  $20 \times e$
- Verification scale interval(e):  $\geq 5 \text{ g}$
- Maximum tare effect:  $\leq -\text{Max}_1$
- Temperature range.  $-10 \text{ }^\circ\text{C}$  to  $+40 \text{ }^\circ\text{C}$
- Power supply: 6 VDC supplied by external 100-240 VAC/DC adapter.

The scale may optional be equipped with an external battery of the type Li-Ion, Pb or Mh. Alternate can dry cells of 1.5VDC be used.

### Model variants and designation

| Model                          | PBII       |            |            |
|--------------------------------|------------|------------|------------|
| Max                            | 15/30 kg   | 30/60 kg   | 60/150 kg  |
| Min                            | 100 g      | 200 g      | 400 g      |
| e =                            | 5/10 g     | 10/20 g    | 20/50 g    |
| T ≤                            | -14.995 kg | -29.990 kg | -59.980 kg |
| E <sub>max</sub> <sup>1)</sup> | 50 kg      | 100 kg     | 250 kg     |

<sup>1)</sup> The load cell fitted in the instrument is a CAS load cell, model BCS-XXPB

### Software

The software is designated yy1.xx,  
yy is an optional county specific code.  
xx is for non-legal minor changes in the software.

This information is displayed at power up.

**Devices**

- Initial zero setting device ( $\leq 20\%$  of Max)
- Semi-automatic zero setting device ( $\leq 4\%$  of Max)
- Zero tracking device ( $\leq 4\%$  of Max)
- Zero indicator
- Net indicator
- Stable weight indicator
- Semi-automatic subtractive tare balancing device
- Gravity compensation
- Weighing unstable samples
- Low battery indicator
- Calibration / set-up mode via sealed internal switch

**Interfaces**

- RS232
- Bluetooth
- USB

