



**OIML Member State**  
Denmark

**OIML Certificate No.**  
**R76/2006-A-DK2-2020.12**

**OIML CERTIFICATE ISSUED UNDER SCHEME A**

**OIML Issuing Authority**

Name: **FORCE Certification A/S**  
Address: Park Allé 345, 2605 Brøndby, Denmark  
Person responsible: Leif Madsen

**Applicant**

Name: **Basculas Prometalicos S.A.**  
Address: Carrera 21 #72-04, Zona Industrial AltaSuiza  
Manizales (Caldas)  
Colombia

**Manufacturer** **Basculas Prometalicos S.A.**

**Identification of the certified type** (*the detailed characteristics will be defined in the additional pages*)

**PRO-KP M / PRO-KC M / PRO-KW M**

**Designation of the module** (*if applicable*)

**Non-automatic electronic weighing instrument**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1, Edition (year): 2006**

For accuracy class (if applicable): **III**

**OIML Certificate No.**  
**R76/2006-A-DK2-2020.12**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type examination report: No. 120-30723.10.20, dated 05 October 2020, that includes 72 pages

Type evaluation report: No. 120-30723.90.20, dated 06 October 2020, that includes 21 pages

The technical documentation relating to the identified type is contained in documentation file:  
No. T20996

**OIML Certificate History**

<b>Revision No.</b>	<b>Date</b>	<b>Description of the modification</b>
Initial version	13 November 2020	

Identification, signature and stamp

**The OIML Issuing Authority**

FORCE Certification A/S

Date: 13 November 2020

Jens Hovgård Jensen

Certification Manager

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

## Descriptive annex

### Characteristics

Type:	PRO-KP M / PRO-KC M / PRO-KW M
Accuracy class:	III
Weighing range:	Single-interval, multi-interval or multi-range
Maximum number of Verification	
Scale Intervals:	$\leq 3000$
Maximum capacity:	3 kg to 30 kg
Minimum capacity:	$20 \times e_i$
Maximum tare effect:	-Max within display limits
Verification scale interval (e=):	$\geq 0.5$ g
Mains power supply:	100-240 VAC, 50/60 Hz using external AC to 10 VDC adapter 6 V internal rechargeable battery (optional)
Operational temperature:	-10 °C to +40 °C

### Software

On PRO-KC M and PRO-KP M can the software version be seen by pressing the T and 5 key simultaneously.

On PRO-KW M can the software version be seen by holding the print key at power up and during the countdown sequence.

The approved software versions are:

PRO-KW M :	100112
PRO-KP M :	200116
PRO-KC M :	300112

### Interfaces

- RS232

### Devices

- Initial zero setting device ( $\leq 20\%$  of Max)
- Semi-automatic zero setting device ( $\leq 4\%$  of Max)
- Zero tracking device ( $\leq 4\%$  of Max)
- Semi-automatic subtractive tare balancing device
- Preset Tare device (PRO-KC M only)
- Gross / Net display (PRO-KW M only)
- Totalization device (PRO-KC M and PRO-KP M. PRO-KP M only when printer is connected)
- Counting device (PRO-KC M and PRO-KW M)
- Printing device
- Gravity compensation device
- Stable equilibrium, Zero, Gross, Net and active range indicators.